

New England Biolabs Certificate of Analysis

Product Name: NEBNext[®] FFPE DNA Repair v2 Module
 Catalog Number: E7360G
 Packaging Lot Number: 10200412
 Expiration Date: 08/2024
 Storage Temperature: -20°C
 Specification Version: PS-E7360S/L/G v1.0

NEBNext [®] FFPE DNA Repair v2 Module Component List			
NEB Part Number	Component Description	Lot Number	Individual QC Result
E7363GVIAL	NEBNext [®] FFPE DNA Repair Buffer v2	10200415	Pass
E7362GVIAL	NEBNext [®] Thermolabile Proteinase K	10200414	Pass
E7361GVIAL	NEBNext [®] FFPE DNA Repair Mix v2	10200413	Pass

Assay Name/Specification	Lot # 10200412
<p>* Individual Product Component Note Standard Quality Control Tests are performed for each component included in NEBNext[®] FFPE DNA Repair v2 Module and meet the designated specifications.</p>	Pass
<p>Functional Testing (NEBNext[®] FFPE DNA Repair v2 Module) Libraries are prepared using 5 and 50 ng of normal human FFPE DNA and 100 ng of human genomic DNA containing oxidative damage using current and previous lots of the NEBNext[®] FFPE DNA Repair v2 Module. Libraries are sequenced on an Illumina[®] sequencing platform and assessed for library yield, sequencing quality metrics, and damaged base repair efficiency.</p>	Pass

This product has been tested and shown to be in compliance with all specifications.

One or more products referenced in this document may be covered by a 3rd-party trademark. Please visit www.neb.com/trademarks for additional information.

Christine Sumner

Christine Sumner
Production Scientist
12 Sep 2023

Josh Hersey

Josh Hersey
Packaging Quality Control Inspector
18 Jan 2024