

## New England Biolabs Certificate of Analysis

**Product Name:** Chitin Magnetic Beads  
**Catalog Number:** E8036S  
**Lot Number:** 10027736  
**Expiration Date:** 11/2021  
**Storage Temperature:** 4°C  
**Specification Version:** PS-E8036S/L v1.0

Chitin Magnetic Beads Component List			
NEB Part Number	Component Description	Lot Number	Individual QC Result
E8036SVIAL	Chitin Magnetic Beads	10027737	Pass

Assay Name/Specification	Lot # 10027736
<b>Binding Capacity (Magnetic Beads)</b> Chitin Magnetic Beads ( 25 µl ) were equilibrated and incubated with 500 µl of crude extract from E. coli containing a plasmid that expresses a SNAP-intein-chitin binding domain fusion protein for 1 hour at 25°C, then washed and the intein cleaved and target eluted. Binding capacity was determined to be >2 mg of target protein per ml of bed volume.	Pass
<b>Functional Binding Assay (Qualitative)</b> Chitin Magnetic Beads ( 25 µl ) were equilibrated and incubated with 500 µl of crude extract from E. coli containing a plasmid that expresses a SNAP-intein-chitin binding domain fusion protein for 1 hour at 25°C, then washed, eluted and evaluated by Tris-Glycine gel to confirm low non-specific binding of extract proteins and high isolation of target.	Pass

This product has been tested and shown to be in compliance with all specifications.



---

Brad Landgraf  
Production Scientist  
29 Nov 2018



---

Michael Tonello  
Packaging Quality Control Inspector  
05 Dec 2018