

240 County Road Ipswich, MA 01938-2723 Tel 978-927-5054 Fax 978-921-1350 www.neb.com info@neb.com

New England Biolabs Certificate of Analysis

Product Name: OneTag® Quick-Load® DNA Polymerase

Catalog Number: M0509L Concentration: 5,000 U/ml

Unit Definition: One unit is defined as the amount of enzyme that will incorporate 15

nmol of dNTP into acid insoluble material in 30 minutes at 75°C.

Packaging Lot Number: 10058631
Expiration Date: 07/2021
Storage Temperature: -20°C

Storage Conditions: 10 mM Tris-HCl, 100 mM KCl, 1 mM DTT, 0.1 mM EDTA, 0.5 % Tween® 20,

0.5 % IGEPAL® CA-630, 50% Glycerol, (pH 7.4 @ 25°C)

Specification Version: PS-M0509S/L/X v1.0

OneTaq® Quick-Load® DNA Polymerase Component List				
NEB Part Number	Component Description	Lot Number	Individual QC Result	
M0509LVIAL	OneTaq® Quick-Load® DNA Polymerase	10047116	Pass	
B9022SVIAL	OneTaq® Standard Reaction Buffer	10034011	Pass	
B0509AVIAL	OneTaq® Quick-Load® Reaction Buffer	10034017	Pass	

Assay Name/Specification	Lot # 10058631
Non-Specific DNase Activity (16 Hour) A 50 μl reaction in NEBuffer 2 containing 1 μg of T3 DNA in addition to a reaction containing Lambda-HindIII DNA and a minimum of 5 units of OneTaq® Quick-Load® DNA Polymerase incubated for 16 hours at 37°C results in a DNA pattern free of detectable nuclease degradation as determined by agarose gel electrophoresis.	Pass
PCR Amplification (5.0 kb Lambda DNA) A 25 μl reaction in OneTaq® Quick-Load® Reaction Buffer in the presence of 200 μM dNTPs and 0.2 μM primers containing 5 ng Lambda DNA with 0.625 units of OneTaq® Quick-Load® DNA Polymerase for 25 cycles of PCR amplification results in the expected 5.0 kb product.	Pass
PCR Amplification (5.0 kb Lambda DNA) A 25 μl reaction in OneTaq® Standard Reaction Buffer in the presence of 200 μM dNTPs and 0.2 μM primers containing 5 ng Lambda DNA with 0.625 units of OneTaq® Quick-Load® DNA Polymerase for 25 cycles of PCR amplification results in the expected 5.0 kb product.	Pass



M0509L / Lot: 10058631

Page 1 of 2

determined by gel electrophoresis using fluorescent detection.

This product has been tested and shown to be in compliance with all specifications.

David Guo

Production Scientist

22 Jul 2019

Michael Tonello

Packaging Quality Control Inspector

03 Dec 2019

