

240 County Road Ipswich, MA 01938-2723 Tel 978-927-5054 Fax 978-921-1350 www.neb.com info@neb.com

New England Biolabs Certificate of Analysis

Product Name: E.coli RNA Polymerase, Core Enzyme

Catalog Number: M0550S Concentration: 1,000 U/ml

Unit Definition: One unit is defined as the amount of enzyme required to incorporate

1 nmole NTP into RNA in 10 minutes at 37°C in the presence of sigma

factor 70.

Packaging Lot Number: 10059647 Expiration Date: 12/2021 Storage Temperature: -20°C

Storage Conditions: 20 mM Tris-HCl, 100 mM NaCl, 1 mM DTT, 0.1 mM EDTA, 50 % Glycerol,

(pH 7.5 @ 25°C)

Specification Version: PS-M0550S v1.0

E.coli RNA Polymerase, Core Enzyme Component List				
NEB Part Number	Component Description	Lot Number	Individual QC Result	
M0550SVIAL	E.coli RNA Polymerase, Core Enzyme	10059646	Pass	
B0550AVIAL	5X E. coli RNA Polymerase Reaction Buffer	10048086	Pass	

Assay Name/Specification	Lot # 10059647
Endonuclease Activity (Nicking) A 50 μl reaction in NEBuffer 4 containing 1 μg of supercoiled PhiX174 DNA and a minimum of 5 units of E. coli RNA Polymerase, Core Enzyme incubated for 4 hours at 37°C results in <10% conversion to the nicked form as determined by agarose gel electrophoresis.	Pass
Exonuclease Activity (Radioactivity Release) A 50 μl reaction in NEBuffer 4 containing 1 μg of a mixture of single and double-stranded [³H] E. coli DNA and a minimum of 5 units of E. coli RNA Polymerase, Core Enzyme incubated for 4 hours at 37°C releases <0.1% of the total radioactivity.	Pass
RNase Activity (Extended Digestion) A 10 µl reaction in NEBuffer 4 containing 40 ng of a 300 base single-stranded RNA and a minimum of 1 unit of E. coli RNA Polymerase, Core Enzyme is incubated at 37°C. After incubation for 4 hours, >90% of the substrate RNA remains intact as determined by gel electrophoresis using fluorescent detection.	Pass

This product has been tested and shown to be in compliance with all specifications.



M0550S / Lot: 10059647 Page 1 of 2



Dongxian Yue Production Scientist 05 Dec 2019

Minichiello

Packaging Quality Control Inspector

07 Jan 2020