

240 County Road Ipswich, MA 01938-2723 Tel 978-927-5054 Fax 978-921-1350 www.neb.com info@neb.com

## New England Biolabs Certificate of Analysis

Product Name: EnGen Lba Cas12a (Cpf1)

Catalog Number: M0653S

Concentration: 1,000 Nanomolar

Lot Number: 10053189
Expiration Date: 08/2021
Storage Temperature: -20°C

Storage Conditions: 500 mM NaCl, 20 mM Sodium Acetate, 0.1 mM EDTA, 0.1 mM TCEP-HCl, 50%

Glycerol, (pH 6.0 @ 25°C)

Specification Version: PS-M0653S v2.0

EnGen Lba Cas12a (Cpf1) Component List				
<b>NEB Part Number</b>	Component Description	Lot Number	Individual QC Result	
M0653SVIAL	ENGEN® LBA CAS12A (CPF1)	10052574	Pass	
B7202SVIAL	NEBuffer™ 2.1	10043906	Pass	

Assay Name/Specification	Lot # 10053189
Non-Specific DNase Activity (16 Hour) A 50 µl reaction in NEBuffer 2.1 containing 1 µg of Lambda DNA and a minimum of 1 pmol of EnGen® Lba Cas12a (Cpf1) incubated for 16 hours at 37°C results in a DNA pattern free of detectable nuclease degradation as determined by agarose gel electrophoresis.	Pass
Functional Testing (Targeted Digestion) A 20 µl reaction in 1X NEBuffer 2.1 containing 20 nM of 100 bp FAM and ROX-labeled double-stranded target DNA, 100 nM crRNA, and 100 nM EnGen® Lba Cas12a (Cpf1) incubated for 15 minutes at 37°C results in ≥90% targeted digestion of the substrate DNA as determined by capillary electrophoresis.	Pass
Exonuclease Activity (Radioactivity Release) A 50 µl reaction in NEBuffer 2.1 containing 1 µg of a mixture of single and double-stranded [ ³H] E. coli DNA and a minimum of 1 pmol of EnGen® Lba Cas12a (Cpf1) incubated for 4 hours at 37°C releases <0.1% of the total radioactivity.	Pass
Endonuclease Activity (Nicking) A 50 μl reaction in NEBuffer 2.1 containing 1 μg of supercoiled PhiX174 RF I DNA and a minimum of 1 pmol of EnGen® Lba Cas12a (Cpf1) incubated for 4 hours at 37°C results in <10% conversion to the nicked form as determined by agarose gel	Pass



M0653S / Lot: 10053189

Page 1 of 2

This product has been tested and shown to be in compliance with all specifications.

Production Scientist

electrophoresis using fluorescent detection.

28 May 2019

Michael Tonello

Packaging Quality Control Inspector

03 Sep 2019

