

New England Biolabs Certificate of Analysis

Product Name: *EcoRI*
Catalog Number: *R0101S*
Concentration: *20,000 U/ml*
Unit Definition: *One unit is defined as the amount of enzyme required to digest 1 µg of Lambda DNA in 1 hour at 37°C in a total reaction volume of 50 µl.*
Lot Number: *10014948*
Expiration Date: *04/2020*
Storage Temperature: *-20°C*
Storage Conditions: *300 mM NaCl, 10 mM KPO4 (pH 7.5), 1 mM DTT, 0.1 mM EDTA, 50% Glycerol, 0.15% Triton X-100, 200 µg/ml BSA*
Specification Version: *PS-R0101S/L v2.0*

EcoRI Component List			
NEB Part Number	Component Description	Lot Number	Individual QC Result
R0101SVIAL	EcoRI	0351804	Pass
B7024SVIAL	Gel Loading Dye, Purple (6X)	10010200	Pass
B0101SVIAL	NEBuffer™ EcoRI	0091805	Pass

Assay Name/Specification	Lot # 10014948
Protein Purity Assay (SDS-PAGE) EcoRI is >95% pure as determined by SDS PAGE analysis using Coomassie Blue detection.	Pass
Exonuclease Activity (Radioactivity Release) A 50 µl reaction in EcoRI Buffer containing 1 µg of a mixture of single and double-stranded [³ H] E. coli DNA and a minimum of 200 units of EcoRI incubated for 4 hours at 37°C releases <0.1% of the total radioactivity.	Pass
Blue-White Screening (Terminal Integrity) A sample of pUC19 vector linearized with a 10-fold excess of EcoRI, religated and transformed into an E. coli strain expressing the LacZ beta fragment gene results in <1% white colonies.	Pass
Ligation and Recutting (Terminal Integrity) After a 20-fold over-digestion of Lambda DNA with EcoRI, >95% of the DNA fragments can be ligated with T4 DNA ligase in 16 hours at 16°C. Of these ligated fragments, >95% can be recut with EcoRI.	Pass

Assay Name/Specification	Lot # 10014948
<p>Non-Specific DNase Activity (16 Hour) A 50 µl reaction in EcoRI Buffer containing 1 µg of Lambda DNA and a minimum of 100 Units of EcoRI incubated for 16 hours at 37°C results in a DNA pattern free of detectable nuclease degradation as determined by agarose gel electrophoresis.</p>	<p>Pass</p>

This product has been tested and shown to be in compliance with all specifications.



Tony Spear-Alfonso
Production Scientist
25 May 2018



Josh Hersey
Packaging Quality Control Inspector
03 Jul 2018