

New England Biolabs Certificate of Analysis

Product Name: BstBI
Catalog Number: R0519S
Concentration: 20,000 U/ml
Unit Definition: One unit is defined as the amount of enzyme required to digest 1 µg of Lambda DNA in 1 hour at 65°C in a total reaction volume of 50 µl.
Packaging Lot Number: 10131691
Expiration Date: 06/2023
Storage Temperature: -20°C
Storage Conditions: 50 mM KCl, 10 mM Tris-HCl (pH 7.4), 1 mM DTT, 0.1 mM EDTA, 50% Glycerol, 200 µg/ml BSA
Specification Version: PS-R0519S/L v1.0

BstBI Component List			
NEB Part Number	Component Description	Lot Number	Individual QC Result
R0519SVIAL	BstBI	10114228	Pass
B6004SVIAL	rCutSmart™ Buffer	10130599	Pass

Assay Name/Specification	Lot # 10131691
<p>Ligation and Recutting (Terminal Integrity) After a 20-fold over-digestion of Lambda DNA with BstBI, >95% of the DNA fragments can be ligated with T4 DNA ligase in 4 hours at 25°C. Of these ligated fragments, >95% can be recut with BstBI.</p>	Pass
<p>Non-Specific DNase Activity (16 Hour) A 50 µl reaction in CutSmart™ Buffer containing 1 µg of Lambda DNA and a minimum of 20 units of BstBI incubated for 16 hours at 65°C results in a DNA pattern free of detectable nuclease degradation as determined by agarose gel electrophoresis.</p>	Pass
<p>Endonuclease Activity (Nicking) A 50 µl reaction in CutSmart™ Buffer containing 1 µg of supercoiled pUC19 DNA and a minimum of 20 units of BstBI incubated for 4 hours at 65°C results in <20% conversion to the nicked form as determined by agarose gel electrophoresis.</p>	Pass
<p>Exonuclease Activity (Radioactivity Release) A 50 µl reaction in CutSmart™ Buffer containing 1 µg of a mixture of single and double-stranded [³H] E. coli DNA and a minimum of 100 units of BstBI incubated for 4 hours at 65°C releases <0.1% of the total radioactivity.</p>	Pass

This product has been tested and shown to be in compliance with all specifications.

One or more products referenced in this document may be covered by a 3rd-party trademark. Please visit www.neb.com/trademarks for additional information.



Penghua Zhang
Production Scientist
17 Dec 2021



Michael Tonello
Packaging Quality Control Inspector
17 Dec 2021