

New England Biolabs Certificate of Analysis


Product Name: *Blpl*
Catalog Number: *R0585S*
Concentration: *10,000 U/ml*
Unit Definition: *One unit is defined as the amount of enzyme required to digest 1 µg of Lambda DNA in 1 hour at 37°C in a total reaction volume of 50 µl.*
Lot Number: *10038688*
Expiration Date: *03/2021*
Storage Temperature: *-20°C*
Storage Conditions: *50 mM NaCl, 10 mM Tris-HCl (pH 7.4), 1 mM DTT, 0.1 mM EDTA, 50% Glycerol, 200 µg/ml BSA*
Specification Version: *PS-R0585S/L v1.0*

Blpl Component List			
NEB Part Number	Component Description	Lot Number	Individual QC Result
R0585SVIAL	Blpl	10038689	Pass
B7204SVIAL	CutSmart® Buffer	10036668	Pass

Assay Name/Specification	Lot # 10038688
Endonuclease Activity (Nicking) A 50 µl reaction in CutSmart™ Buffer containing 1 µg of supercoiled PhiX174 DNA and a minimum of 30 units of Blpl incubated for 4 hours at 37°C results in <20% conversion to the nicked form as determined by agarose gel electrophoresis.	Pass
Exonuclease Activity (Radioactivity Release) A 50 µl reaction in CutSmart™ Buffer containing 1 µg of a mixture of single and double-stranded [³ H] E. coli DNA and a minimum of 100 units of Blpl incubated for 4 hours at 37°C releases <0.1% of the total radioactivity.	Pass
Ligation and Recutting (Terminal Integrity) After a 10-fold over-digestion of Lambda DNA with Blpl, >95% of the DNA fragments can be ligated with T4 DNA ligase in 16 hours at 16°C. Of these ligated fragments, ~75% can be recut with Blpl.	Pass
Non-Specific DNase Activity (16 Hour) A 50 µl reaction in CutSmart™ Buffer containing 1 µg of Lambda DNA and a minimum of 10 units of Blpl incubated for 16 hours at 37°C results in a DNA pattern free of detectable nuclease degradation as determined by agarose gel electrophoresis.	Pass

Assay Name/Specification	Lot # 10038688
Protein Purity Assay (SDS-PAGE) B1p1 is >95% pure as determined by SDS PAGE analysis using Coomassie Blue detection.	Pass

This product has been tested and shown to be in compliance with all specifications.



Anthony Francis
Production Scientist
06 Mar 2019



Michael Tonello
Packaging Quality Control Inspector
26 Apr 2019