

New England Biolabs Certificate of Analysis

Product Name: EcoRI-HF[®]

Catalog #: R3101S/L

Concentration: 20,000 units/ml

Unit Definition: One unit is defined as the amount of enzyme required to digest 1 µg of Lambda DNA in 1 hour at 37°C in a total reaction volume of 50 µl.

Lot #: 0091408

Assay Date: 08/2014

Expiration Date: 8/2016

Storage Temp: -20 °C

Storage Conditions: 300 mM NaCl, 10 mM KPO₄ (pH 7.5), 10 mM DTT, 0.1 mM EDTA, 50% Glycerol, 0.15% Triton X-100, 200 µg/ml BSA

Specification Version: PS-R3101S/L v1.0

Effective Date: 06 Aug 2013

Assay Name/Specification (minimum release criteria)	Lot #0091408
Blue-White Screening (Terminal Integrity) - A sample of pUC19 vector linearized with a 10-fold excess of EcoRI-HF [™] , religated and transformed into an <i>E. coli</i> strain expressing the LacZ beta fragment gene results in <1% white colonies.	Pass
Endonuclease Activity (Nicking) - A 50 µl reaction in CutSmart [™] Buffer containing 1 µg of supercoiled PhiX174 DNA and a minimum of 100 Units of EcoRI-HF [™] incubated for 4 hours at 37°C results in <10% conversion to the nicked form as determined by agarose gel electrophoresis.	Pass
Exonuclease Activity (Radioactivity Release) - A 50 µl reaction in CutSmart [™] Buffer containing 1 µg of a mixture of single and double-stranded [³ H] <i>E. coli</i> DNA and a minimum of 200 units of EcoRI-HF [™] incubated for 4 hours at 37°C releases <0.1% of the total radioactivity.	Pass
Ligation and Recutting (Terminal Integrity) - After a 20-fold over-digestion of Lambda DNA with EcoRI-HF [™] , >95% of the DNA fragments can be ligated with T4 DNA ligase in 16 hours at 16°C. Of these ligated fragments, >95% can be recut with EcoRI-HF [™] .	Pass
Non-Specific DNase Activity (16 Hour) - A 50 µl reaction in CutSmart [™] Buffer containing 1 µg of Lambda DNA and a minimum of 100 Units of EcoRI-HF [™] incubated for 16 hours at 37°C results in a DNA pattern free of detectable nuclease degradation as determined by agarose gel electrophoresis.	Pass
Protein Purity Assay (SDS-PAGE) - EcoRI-HF [™] is >95% pure as determined by SDS PAGE analysis using Coomassie Blue detection.	Pass





240 County Road
Ipswich, MA 01938-2723

Tel 978-927-5054
Fax 978-921-1350

www.neb.com
info@neb.com

New England Biolabs Certificate of Analysis

* The BSA in this product has been granted an EDQM "Certificate of Suitability" from the European Directorate for the Quality of Medicines (# R1-CEP-2003-204-Rev00) and has been granted a USDA Certificate for Export of Bovine Blood Plasma/Serum for Manufacture into Pharmaceutical Products.

Authorized by
Derek Robinson
06 Aug 2013

Inspected by
Casey Madinger
26 Aug 2014

