

New England Biolabs Certificate of Analysis

Product Name: *KpnI-HF*[®]
Catalog #: R3142S/L
Concentration: 20,000 units/ml
Unit Definition: One unit is defined as the amount of enzyme required to digest 1 µg of pXba DNA in 1 hour at 37°C in a total reaction volume of 50 µl.
Lot #: 0071709
Assay Date: 09/2017
Expiration Date: 9/2019
Storage Temp: -20°C
Storage Conditions: 50 mM KCl, 10 mM Tris-HCl (pH 7.4), 1 mM DTT, 0.1 mM EDTA, 50% Glycerol, 200 µg/ml BSA
Specification Version: PS-R3142S/L v1.0
Effective Date: 20 May 2015

Assay Name/Specification (minimum release criteria)	Lot #0071709
Blue-White Screening (Terminal Integrity) - A sample of Litmus28i vector linearized with a 10-fold excess of KpnI-HF [™] , religated and transformed into an <i>E. coli</i> strain expressing the LacZ beta fragment gene results in <1% white colonies.	Pass
Endonuclease Activity (Nicking) - A 50 µl reaction in CutSmart [™] Buffer containing 1 µg of supercoiled PhiX174 DNA and a minimum of 100 Units of KpnI-HF [™] incubated for 4 hours at 37°C results in <10% conversion to the nicked form as determined by agarose gel electrophoresis.	Pass
Exonuclease Activity (Radioactivity Release) - A 50 µl reaction in CutSmart [™] Buffer containing 1 µg of a mixture of single and double-stranded [³ H] <i>E. coli</i> DNA and a minimum of 200 units of KpnI-HF [™] incubated for 4 hours at 37°C releases <0.1% of the total radioactivity.	Pass
Ligation and Recutting (Terminal Integrity) - After a 50-fold over-digestion of pXba DNA with KpnI-HF [™] , >95% of the DNA fragments can be ligated with T4 DNA ligase in 16 hours at 16°C. Of these ligated fragments, >95% can be recut with KpnI-HF [™] .	Pass
Non-Specific DNase Activity (16 Hour) - A 50 µl reaction in CutSmart [™] Buffer containing 1 µg of pXba DNA and a minimum of 100 Units of KpnI-HF [™] incubated for 16 hours at 37°C results in a DNA pattern free of detectable nuclease degradation as determined by agarose gel electrophoresis.	Pass
Protein Purity Assay (SDS-PAGE) - KpnI-HF [™] is >95% pure as determined by SDS PAGE analysis using Coomassie Blue detection.	Pass

New England Biolabs Certificate of Analysis

* The BSA in this product has been granted an EDQM "Certificate of Suitability" from the European Directorate for the Quality of Medicines (# R1-CEP-2003-204-Rev00) and has been granted a USDA Certificate for Export of Bovine Blood Plasma/Serum for Manufacture into Pharmaceutical Products.



Authorized by
Derek Robinson
20 May 2015



Inspected by
Jianying Luo
26 Sep 2017

