

New England Biolabs Certificate of Analysis

Product Name: *MluI-HF*
Catalog #: *R3198S/L*
Concentration: *20,000 units/ml*
Unit Definition: *One unit is defined as the amount of enzyme required to digest 1 µg of Lambda DNA in 1 hour at 37°C in a total reaction volume of 50 µl.*
Lot #: *0011508*
Assay Date: *08/2015*
Expiration Date: *8/2017*
Storage Temp: *-20°C*
Storage Conditions: *200 mM NaCl, 10 mM Tris-HCl (pH 7.4), 1 mM DTT, 0.1 mM EDTA, 50 % Glycerol, 200 µg/ml BSA*
Specification Version: *PS-R3198S/L v1.0*
Effective Date: *20 Nov 2014*


Assay Name/Specification (minimum release criteria)	Lot #0011508
Endonuclease Activity (Nicking) - A 50 µl reaction in CutSmart™ Buffer containing 1 µg of supercoiled pUC19 DNA and a minimum of 60 units of MluI-HF incubated for 4 hours at 37°C results in <10% conversion to the nicked form as determined by agarose gel electrophoresis.	Pass
Exonuclease Activity (Radioactivity Release) - A 50 µl reaction in CutSmart™ Buffer containing 1 µg of a mixture of single and double-stranded [³ H] <i>E. coli</i> DNA and a minimum of 100 units of MluI-HF incubated for 4 hours at 37°C releases <0.1% of the total radioactivity.	Pass
Functional Test (15 minute Digest) - A 50 µl reaction in CutSmart™ Buffer containing 1 µg of Lambda DNA and 1 µl of MluI-HF incubated for 15 minutes at 37°C results in complete digestion as determined by agarose gel electrophoresis.	Pass
Ligation and Recutting (Terminal Integrity) - After a 20-fold over-digestion of Lambda DNA with MluI-HF, >95% of the DNA fragments can be ligated with T4 DNA ligase in 16 hours at 16°C. Of these ligated fragments, >95% can be recut with MluI-HF.	Pass
Non-Specific DNase Activity (16 Hour) - A 50 µl reaction in CutSmart™ Buffer containing 1 µg of Lambda DNA and a minimum of 100 units of MluI-HF incubated for 16 hours at 37°C results in a DNA pattern free of detectable nuclease degradation as determined by agarose gel electrophoresis.	Pass
Protein Purity Assay (SDS-PAGE) - MluI-HF is ≥ 95% pure as determined by SDS-PAGE analysis using Coomassie Blue detection.	Pass

New England Biolabs Certificate of Analysis

** The BSA in this product has been granted an EDQM "Certificate of Suitability" from the European Directorate for the Quality of Medicines (# R1-CEP-2003-204-Rev00) and has been granted a USDA Certificate for Export of Bovine Blood Plasma/Serum for Manufacture into Pharmaceutical Products.*



Authorized by
Derek Robinson
20 Nov 2014



Inspected by
Timothy Meixsell
25 Aug 2015

