

## New England Biolabs Certificate of Analysis

**Product Name:** *Monarch® RNA Purification Columns*  
**Catalog Number:** *T2007L*  
**Packaging Lot Number:** *10210907*  
**Expiration Date:** *11/2025*  
**Storage Temperature:** *25°C*  
**Specification Version:** *PS-T2007L v2.0*

Monarch® RNA Purification Columns Component List			
NEB Part Number	Component Description	Lot Number	Individual QC Result
T2118-1	Monarch® Spin Collection Tubes	10234064	<b>Pass</b>
T2007-1	Monarch® RNA Purification Columns	10239817	<b>Pass</b>

Assay Name/Specification	Lot # 10210907
<b>Functional Testing (DNA Recovery and Purity)</b> Twenty-four Monarch® RNA Purification Columns are tested with 10 µg of Lambda/BstEII Marker resulting in ≥75% recovery in ≥80% of the samples. OD 260/280 and 260/230 are ≥1.75 in ≥80% of the samples.	<b>Pass</b>

This product has been tested and shown to be in compliance with all specifications.

One or more products referenced in this document may be covered by a 3rd-party trademark. Please visit [www.neb.com/trademarks](http://www.neb.com/trademarks) for additional information.



Julie Gaumont  
Production Scientist  
02 May 2024



Julie Gaumont  
Packaging Quality Control Inspector  
02 May 2024