


New England Biolabs Certificate of Analysis

Product Name: *Monarch RNA Cleanup Columns (10 µg)*
Catalog Number: *T2037L*
Packaging Lot Number: *10059913*
Expiration Date: *08/2021*
Storage Temperature: *25°C*
Specification Version: *PS-T2037L v1.0*

Assay Name/Specification	Lot # 10059913
Functional Testing (RNA Recovery and Purity) Twenty-four Monarch® RNA Cleanup Columns (T2037, 10 µg) are tested with 1 µg of rRNA resulting in ≥75% recovery in ≥80% of the samples. OD 260/280 and 260/230 are ≥1.75 in ≥80% of the samples.	Pass

This product has been tested and shown to be in compliance with all specifications.



Ana Egana
 Production Scientist
 26 Mar 2020

Darcie Spaulding
 Packaging Quality Control Inspector
 26 Mar 2020