


New England Biolabs Certificate of Analysis

Product Name: *Monarch gDNA Purification Columns*
 Catalog Number: *T3017L*
 Packaging Lot Number: *10041709*
 Expiration Date: *02/2021*
 Storage Temperature: *25°C*
 Specification Version: *PS-T3017L v1.0*

Assay Name/Specification	Lot # 10041709
Functional Testing (DNA Recovery and Purity) Twenty-four Monarch® gDNA Purification Columns are tested with 5 µg of HindIII digested Lambda resulting in ≥75% recovery in ≥80% of the samples. OD 260/280 and 260/230 are ≥1.75 in ≥80% of the samples.	Pass

This product has been tested and shown to be in compliance with all specifications.



Ana Egana
Production Scientist
25 Nov 2019

Darcie Spaulding
Packaging Quality Control Inspector
25 Nov 2019