

New England Biolabs Product Specification

<i>Product Name:</i>	<i>Monarch[®] HMW DNA Extraction Kit for Tissue</i>
<i>Catalog #:</i>	<i>T3060G</i>
<i>Kit Components:</i>	<i>Monarch[®] DNA Capture Beads (T3005) — Store at 25°C</i> <i>Monarch[®] Bead Retainers (T3004) — Store at 25°C</i> <i>Monarch[®] 2 ml Tubes (T3003) — Store at 25°C</i> <i>Monarch[®] Pestles (T3002) — Store at 25°C</i> <i>Monarch[®] Pestle Tubes (T3001) — Store at 25°C</i> <i>Monarch[®] Spin Collection Tubes (T2118) — Store at 25°C</i> <i>Monarch[®] HMW gDNA Tissue Lysis Buffer (T3061) — Store at 25°C</i> <i>Monarch[®] Protein Separation Solution (T3062) — Store at 25°C</i> <i>Monarch[®] gDNA Wash Buffer (T3015) — Store at 25°C</i> <i>Monarch[®] gDNA Elution Buffer II (T3056) — Store at 25°C</i> <i>Monarch[®] RNase A (T3018) — Store at -20°C</i> <i>Proteinase K, Molecular Biology Grade (P8200) — Store at 25°C</i>
<i>Shelf Life:</i>	<i>24 months</i>
<i>Storage Temp:</i>	<i>Multi-temperature</i>
<i>Specification Version:</i>	<i>PS-T3060G v3.0</i>
<i>Effective Date:</i>	<i>18 Mar 2024</i>

Assay Name/Specification (minimum release criteria)

Functional Testing (Genomic DNA Yield and Integrity Analysis) (Tissue) - Genomic DNA is purified from 8 samples of NEB 10-beta Competent *E. coli* using the High Molecular Weight DNA Extraction from Bacteria protocol. Each set of reagents is functionally tested to ensure successful isolation of gDNA from 1 x 10e9 cells. Yield of genomic DNA is ≥6 µg in ≥80% of the samples. OD 260/280 and 260/230 are ≥1.75 in ≥80% of the samples, and DIN values are ≥9 in ≥80% of the samples. In ≥ 80% of samples, ≥ 75% of the DNA is 50 kb in size or greater.

* **Individual Product Component Note** - Standard Quality Control Tests are performed for each component included in Monarch[®] HMW DNA Extraction Kit for Tissue and meet the designated specifications.





240 County Road
Ipswich, MA 01938-2723

Tel 978-927-5054
Fax 978-921-1350

www.neb.com
info@neb.com

New England Biolabs Product Specification

*One or more products referenced in this document may be covered by a 3rd-party trademark.
Please visit www.neb.com/trademarks for additional information.*

Date 18 Mar 2024

Lauren Brown
Quality Approver

